



Dental Inspection, Maintenance, and Waste Disposal Log



Examples:

Date	Person	Vendor / Equipment	Result / Action Taken / Notes
2/10/2020	John Smith	Dental R US	Checked pump, filter, no leaks, retaining unit cracked & replaced.
3/15/2020	Jane Doe	Amalgam Waste Bucket	Began use. Wrote date on recycle container. Model #ABC123
6/30/2020	Bob B.	Unplug-It LLC System maintenance	Dirty, leaking. Replaced filter and fixed leak. Container is ½ full.
10/5/2020	John Smith	Waste amalgam (UPS/Fed Ex/etc)	Shipped off-site. ⁽¹⁾ Replaced container.

(1) Per 40 CFR 441.5(b), you must have documentation of all dates that collected dental amalgam is picked up or shipped for proper disposal. Make sure the treatment, storage, or disposal facility receiving the amalgam retaining container(s) is permitted or licensed. The documentation must be maintained and available for inspection in either physical or electronic form for a minimum of three years.